

**Weightron Bilanciai Ltd,**  
Titan Works  
Bridge Way off Broombank Road,  
Chesterfield Trading Estate,  
Chesterfield S41 9QJ, UK  
Tel: 01246 260062  
Fax: 01246 260844  
www.weightron.com



**Weightron Bilanciai Ltd,**  
Unit 4b, Block 4  
Westpark, Chelston  
Wellington TA21 9AD, UK  
Tel: 01823 662355  
Fax: 01823 662017  
www.weightron.com



Members of the  
UK Weighing Federation  
Membership No. 1070

## CALIBRATION CERTIFICATE

This Calibration Certificate is issued in accordance with the UKWF Code of Practice for the calibration of Non Automatic Weighing Machines. All weights used are traceable to NPL

**Customer**

NICKS LTD
SCARTH ROAD
SOWERY INB EST
BARRON IN FURNACE
LOCATION LA14 4QR

**Machine Details**

Cust ref	
Make	BILANCIAI
Model	D 800
Ser No	11814
Capacity	15000 KG
e/d	120 / KG
Working cap	

**Test Weights**

Set No.	Cert No.	Date Cal
3	TE10302	DEC 17

  

Cert Interval	12 MONTHS
Next Cert	JULY 2019

**Tolerance +/-**

0-5000	1e
501-20000	2e
>20000	3e

**Comments** WEATHER DAMP & CALM  
FORK LIFT 3920

2	4	6
1	3	5

**AS FOUND LINEARITY & HYSTERESIS**

Load	Loading	Dev	Unloading	Dev
0	0	0	0	0
10000	10000	0	10000	0
20000	19980	-20	19980	-20
LOAD ✓	19100			
29900	29900	0	29900	0
39900	39900	0	39900	0
43840	43840	0	43840	0

**DEFINITIVE LINEARITY & HYSTERESIS**

Load	Loading	Dev	Unloading	Dev

**Repeatability**

Load	Ind	Ind	Ind
2000	2000	2000	2000
43840	43840	43840	43840

**Repeatability**

Load	Ind	Ind	Ind

**Eccentric Load**

Reading	Dev	Reading	Dev
1 10000	0	2 5940	20
3 9980	-20	4 9980	-20
5 10000	0	6 9980	-20
7		8	

**Eccentric Load**

Reading	Dev	Reading	Dev
1		2	
3		4	
5		6	
7		8	

Signature H Wood Name H Wood Date 13-7-18

**Weightron Bilanciai Ltd,**  
Titan Works  
Bridge Way off Broombank Road,  
Chesterfield Trading Estate,  
Chesterfield S41 9QJ, UK  
Tel: 01246 260062  
Fax: 01246 260844  
www.weightron.com



**Weightron Bilanciai Ltd,**  
Unit 4b, Block 4  
Westpark, Chelston  
Wellington TA21 9AD, UK  
Tel: 01823 662355  
Fax: 01823 662017  
www.weightron.com



Members of the  
UK Weighing Federation  
Membership No. 1070

## CALIBRATION CERTIFICATE

*This Calibration Certificate is issued in accordance with the UKWF Code of Practice for the calibration of Non Automatic Weighing Machines. All weights used are traceable to NPL*

**Customer**

HICKS LTD SCARTH ROAD SOWERBY IND ESTATE BARROW IN FURNACE LA14 4QR LOCATION
---

**Machine Details**

Cust ref	
Make	WORLD WIDE
Model	E302-M
Ser No	P149T071
Capacity	13000 1KG
e/d	1 1 1KG
Working cap	

**Test Weights**

Set No.	Cert No.	Date Cal
3	TR1032	DEC 17

  

Cert Interval	12 MONTHS
Next Cert	JULY 2019

Tolerance +/-	Weight
	0 = 1 KG
0-500g	1 KG
500-2000g	2 KG
2000g	3 KG

Comments ... End ... MIDDLE END ... 1000 KG	<table border="1" style="margin: auto;"> <tr><td style="width: 30px;">2</td><td style="width: 30px;">3</td><td style="width: 30px;">4</td></tr> <tr><td> </td><td style="text-align: center;">2</td><td> </td></tr> <tr><td> </td><td style="text-align: center;">1</td><td style="text-align: center;">3</td></tr> </table> INS	2	3	4		2			1	3
2	3	4								
	2									
	1	3								

**AS FOUND LINEARITY & HYSTERESIS**

Load	Loading	Dev	Unloading	Dev
0	0	0	0	0
1000	1000	0	1000	0
2000	2000	0	2000	0
3000	3000	1	3000	1

**DEFINITIVE LINEARITY & HYSTERESIS**

Load	Loading	Dev	Unloading	Dev

**Repeatability**

Load	Ind	Ind	Ind
1000	1000	1000	1000

**Repeatability**

Load	Ind	Ind	Ind

**Eccentric Load**

Reading	Dev	Reading	Dev
1 1000	0	2 1000	0
3 1000	0	4	
5		6	
7		8	

**Eccentric Load**

Reading	Dev	Reading	Dev
1		2	
3		4	
5		6	
7		8	

Signature H. J. Wood Name H. Wood Date 13-7-18