

Weightron Bilanciai Ltd,
Titan Works
Bridge Way off Broombank Road,
Chesterfield Trading Estate,
Chesterfield S41 9QJ, UK
Tel: 01246 260062
Fax: 01246 260844
www.weightron.com



Weightron Bilanciai Ltd,
Unit 4b, Block 4
Westpark, Chelston
Wellington TA21 9AD, UK
Tel: 01823 662355
Fax: 01823 662017
www.weightron.com



CALIBRATION CERTIFICATE

This Calibration Certificate is issued in accordance with the UKWF Code of Practice for the calibration of Non Automatic Weighing Machines. All weights used are traceable to NPL

Customer

Wicks
SCARTH ROAD
SOWERBY IND EST
BARROW-IN-FURNESS
CUMBRIA
LOCATION LA14 4QR.

Machine Details

Cust ref	WEIGHBRIDGE
Make	SBP.M
Model	D300
Ser No	113142
Capacity	50 1000 /kg
e/d	1 20 /kg
Working cap	

Test Weights

Set No.	Cert No.	Date Cal
T11-20	7854	15-10-19

Cert Interval	YEARLY	
Next Cert	APRIL	2021

Tolerance +/-	Kg
0-10000	20
10000-40000	40
40000-50000	60

Comments

2	4	6
1	3	5

AS FOUND LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev
2000	2000	0	0	0
8000	8000	0	2000	0
12000	12000	0	8000	0
16000	16000	0	12000	0
20000	20000	0	16000	0
20400	20400	0	20400	0
30400	30400	0	30400	0
40400	40400	0	40400	0
45020	45040	20	45040	20

DEFINITIVE LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev

Repeatability

Load	Ind	Ind	Ind
45020	45040	45020	45040

Repeatability

Load	Ind	Ind	Ind

Eccentric Load 6620kg

Reading	Dev	Reading	Dev
1 6620	0	2 6640	20
3 6620	0	4 6620	0
5 6620	0	6 6620	0
7		8	

Eccentric Load

Reading	Dev	Reading	Dev
1		2	
3		4	
5		6	
7		8	

Signature Name D. Waller Date 28-4-2020

C1.3.88.1M

Weightron Bilanciai Ltd,
Titan Works
Bridge Way off Broombank Road,
Chesterfield Trading Estate,
Chesterfield S41 9QJ, UK
Tel: 01246 260062
Fax: 01246 260844
www.weightron.com



Weightron Bilanciai Ltd,
Unit 4b, Block 4
Westpark, Chelston
Wellington TA21 9AD, UK
Tel: 01823 662355
Fax: 01823 662017
www.weightron.com



CALIBRATION CERTIFICATE

This Calibration Certificate is issued in accordance with the UKWF Code of Practice for the calibration of Non Automatic Weighing Machines. All weights used are traceable to NPL

Customer

WICKS SCARTH ROAD SOWERBY IND EST BARROW-IN-FURNESS CUMBRIA
LOCATION LA14 4QR SITE

Machine Details

Cust ref	PLATFORM
Make	PRM
Model	E302M
Ser No	PL49T071
Capacity	31000 kg
e/d	1 i / kg
Working cap	

Test Weights

Set No.	Cert No.	Date Cal
T11-1-22	7854	15-10-19

Cert Interval	YEARLY
Next Cert	APRIL 2021

Tolerance +/-	kg
0-500	1
500-2000	2
2000-3000	3

Comments

2	3
1	4

AS FOUND LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev
400	400	0	0	0
1000	1000	0	400	0
2000	2000	0	1000	0
3000	3000	0	2000	0

DEFINITIVE LINEARITY & HYSTERESIS

Load	Loading	Dev	Unloading	Dev

Repeatability 2000kg

Load	Ind	Ind	Ind
2000	2000	2001	2000

Repeatability

Load	Ind	Ind	Ind

Eccentric Load 400kg

Reading	Dev	Reading	Dev
1 400	0	2 400	0
3 400	0	4 400	0
5		6	
7		8	

Eccentric Load

Reading	Dev	Reading	Dev
1		2	
3		4	
5		6	
7		8	

Signature Name D. Walker Date 28-4-2020

C1.3.88.1M